



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/775,666	02/09/2004	Ramez Emile Necola Shehada	064693-0092	1472

7590 06/26/2007
MCDERMOTT, WILL & EMERY
Suite 3400
2049 Century Park East
Los Angeles, CA 90067

EXAMINER

HILL, LAURA C

ART UNIT	PAPER NUMBER
----------	--------------

3761

MAIL DATE	DELIVERY MODE
-----------	---------------

06/26/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.



UNITED STATES DEPARTMENT OF COMMERCE
U.S. Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450

APPLICATION NO./ CONTROL NO.	FILING DATE	FIRST NAMED INVENTOR / PATENT IN REEXAMINATION	ATTORNEY DOCKET NO.
---------------------------------	-------------	---	---------------------

EXAMINER

ART UNIT	PAPER
----------	-------

20070301

DATE MAILED:

Please find below and/or attached an Office communication concerning this application or proceeding.

Commissioner for Patents

This SUPPLEMENTAL Examiner's Amendment contains the IDS dated 26 April 2007 filed after the original Examiner's Amendment dated 9 March 2007 but prior to payment of the issue fee on 25 May 2007. The IDS has been considered herein. No action is required by Applicant.

Supplemental
Notice of Allowability

Application No.

10/775,666

Examiner

Laura C. Hill

Applicant(s)

NECOLA SHEHADA ET AL.

Art Unit

3761

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to Information Disclosure Statement filed 26 April 2007.
2. ☒ The allowed claim(s) is/are 68-79.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date 4/26/07
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

SUPPLEMENTAL EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Marc Brown on 1 March 2007.

Please **change the title to:**

Implanted Surgical Drain with Sensing and Transmitting Elements for Monitoring Internal Tissue Condition.

Reasons for Allowance

Information Disclosure Statement

The information disclosure statement (IDS) submitted on 26 April 2007 was filed after the mailing date of the Examiner's Amendment on 9 March 2007. The submission is in compliance with the provisions of 37 CFR 1.97. Accordingly, the information disclosure statement is being considered by the examiner.

Claims 68-79 are allowed.

The following is an examiner's statement of reasons for allowance: the art of record alone or in combination fails to disclose or fairly suggest a surgical drain configured to rest against the surface of the tissue and configured to not penetrate the tissue that has transmitting & draining elements integrated with a first outer surface of

Art Unit: 3761

the drain, in combination with a processor that determines a first color that is representative of detected spectral energy from a sensing element, and a display configured to depict the color to assist a physician in determining the health of at least one tissue. The following is the closest prior art:

Benaron et al. (US 5,769,791) discloses surgical tool 30,210 with tip 40 passing through or around internal body tissue (column 10, lines 29-38) or alternatively resting on tissue 207 and not penetrating through ti (column 18, lines 14-15, figure 6) comprising first transmitting element 22, 241,242 provides a light control signal using optical fibers (column 9, lines 23-31, column 18, lines 15-18), first sensing element 24, 245, 246 that receives signals corresponding to the detected light (column 18, lines 20-21, figure 6), performs a color analysis and ascertains the desired spectral characteristics of the detected light, and provides the spectral data to the signal processor 26 (column 9, lines 32-36) which in turn provides the type of tissue to display system 28 (column 9, lines 38-42). However, Benaron doesn't disclose a surgical drain having a tube to transport drained fluid out of the body or sensor and transmitter embedded within tool.

Crowley (US 6,882,875) discloses an interventional device 10, 90 inserted into the body (column 3, lines 15-16) and positioned against tissue 92 (column 5, lines 1-2, figure 5A) comprising a first transmitter 20 on the inner surface of the body (column 3, lines 25-40, figure 1), a sensor or light source 16 or LEDs 38a, 38b for providing a signal (column 3, lines 45-52, column 4, lines 16-20), processor and display 14 (column 3, lines 19-24) but no surgical drain having a tube to transport drained fluid out of the

Art Unit: 3761

body or sensor and transmitter embedded within tool or motivation to combine with surgical drains since all other surgical drains (e.g. Russo US 4,317,452) penetrate inside the tissue and don't have sensing or transmitting means.

Mayevsky (US 5,916,171) discloses a device that simultaneously monitors NADH, CBF and other physiological properties using a bundle of optical fibers resting against the surface of the tissue (column 4, lines 10-20, column 6, lines 52-53, columns 9-10, table 3) but no surgical drain.

Frank (US 3,769,497) discloses a biomedical device having a surgical drainage tubing 13 coming from a catheterized patient having transmitting switch 19 and sensing device 16 for transmitting and detecting energy (column 2, lines 7-38, figure 1). One would not be motivated to modify the sensing devices of Benaron, Crowley or Mayevsky listed above with the drain since Frank discloses a drain that *penetrates* the tissue. Even if the modification took place, the transmitting and sensing devices 19, 16 respectively are not integrated on the outer surface of the drain but rather are located remotely from the drain 13 (figure 1).

Takezawa et al. (US 5,108,364) discloses catheter 1 that is *implanted into a body cavity* and thus penetrates the tissue (column 3, lines 55-65) having temperature sensor 10 that detects energy and embedded within the catheter (column 4, lines 9-15), processor and display (column 4, lines 9-10) and drainage holes 5 (column 3, lines 29-38, figure 1A), and a second sensor that can simultaneously measure temperature at a plurality of regions (column 4, lines 24-26). However, there is no transmitter or colors that are determined or depicted but rather temperature is monitored.

Bedingham (US 5,421,328) discloses an arterial catheter 53 with oxygen sensor 69, carbon dioxide sensor 71, and pH sensor 73 affixed to a distal end of transmitting element/optical fibers 75,77,79 on an inner surface of the catheter (column 7, lines 46-51, column 8, lines 2-4 and figure 2). Thus the catheter is implanted and penetrates the tissue.

Sullivan et al. (US 4,497,324) discloses a urinary , interperitoneal, intercardial, respiratory or intervacular catheter penetrating the tissue (column 5, lines 7-11) having drainage lumen 20 (column 3, lines 31-39), temperature transducer/sensor 32 imbedded in the catheter wall to detect temperature and convert said temperature to an electrical signal (column 3, lines 60-63), and a transmitting element/electrical lead 34 imbedded in the catheter wall (column 4, lines 19-20).

Yanda (US 4,413,633) discloses catheter tube 10 inserted into a urethra and thus penetrating the tissue that forms the urethra and connected to drainage tube 30 (column 2, lines 32-35 and lines 42-47), sensor 38 and transmitter on inner surface of the catheter (column 3, lines 18-39 and figure 2).

Pavoni et al. (US 5,906,584) discloses electrodes/sensors 5 on outer surface of catheter 3 but catheter is inserted within tissue, and a monitor/reading apparatus and a power supply/energy delivery transmitting element (column 4, line 36-column 5, line 8, figures 6-7). Also there is no drain disclosed.

Frazer (WO 92/11803) discloses cardiopulmonary monitoring system 100 that transmits and receives optical signals to and from the interior of the penetrated blood vessel via fiber optic catheter 10, monitor/display, and processing (see abstract, etc).

Art Unit: 3761

Hall et al. (US 6,210,346) discloses a catheter 54 with sensor 76, drainage holes, fiber optic cable 98 but the catheter is inserted *into* the ventricular and/or parenchymal spaces of the brain and thus *penetrates the tissue*.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Laura C. Hill whose telephone number is 571-272-7137. The examiner can normally be reached on Monday through Friday (hours vary).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tatyana Zalukaeva can be reached on 571-272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3761

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Laura C. Hill
Examiner
Art Unit 3761
LCH



TATYANA ZALUKAEVA
SUPERVISORY PRIMARY EXAMINER